



8100 National Dr. - Little Rock, AR 72209
501-455-3233 Fax 501-455-6118

24 September 2025

Camille Gernhart
OSG
5301 Northshore Dr
North Little Rock, AR 72118

Project: Samples
Project Number: September 2025 -- EVT
SDG Number: 2509365

Enclosed are the results of analyses for samples received by the laboratory on 16-Sep-25 14:05. If you have any questions concerning this report, please feel free to contact me.

Sample Receipt Information:

Custody Seals	✓
Containers Correct	✓
COC/Labels Agree	✓
Received On Ice	✓
Temperature on Receipt	3.0°C

Sincerely,

A handwritten signature in blue ink that reads "Norma James".

Norma James
Technical Director

This document is intended only for the use of the person(s) to whom it is expressly addressed. This document may contain information that is confidential and legally privileged. If you are not the intended recipient, you are notified that any disclosure, distribution, or copying of this document is strictly prohibited. If you have received this document in error, please destroy.

24 September 2025

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05



CASE NARRATIVE

Sample Delivery Group – 2509365

One OR more of the qualifiers described below may appear in this report. Qualifiers in **RED apply to this SDG (Sample Delivery Group).**

CALIBRATION QUALIFIERS:

<u>Qualifier</u>	<u>Description</u>
CR	Result above highest calibration standard, but within linear calibration range.
Est3	Result at the instrument was above the concentration of the highest standard in the calibration curve.
E2-F	Second Source Verification Failure
E7	Internal Standard Response Failure
E11	Initial Calibration Minimum Response Factor Failure
E21	CCV Low
E-01	CCV High
E35	Low Level CCV Failure

QUALITY CONTROL QUALIFIERS:

<u>Qualifier</u>	<u>Description</u>
E20	Sample used as "parent" for the associated analytical batch.
%D3/S-01	Surrogate failed to recover within acceptance criteria (%D3/S-01).
E1	Results associated with this surrogate were qualified as "estimated" (E1).
B	Present in the Associated Blank
B1	Present in Blank, but Not In the Sample.
%D2 / E5/NREC	Laboratory Control Spike (LCS) and/or Laboratory Control Spike Duplicate (LCSD) failed to recover with acceptance criteria (%D2). Associated results were qualified as "estimated" (E5 and/or NREC (No Recovery)).
%D1	Matrix Spike (MS) and/or Matrix Spike Duplicate (MSD) failed acceptance criteria.
MBA	Failed criteria due to the high concentration of analyte in the parent sample.
MBI	Failed criteria due to an interference in the parent sample.
%D3	Quality Control Surrogate failed acceptance criteria.
NREC	Quality Control failed: No Recovery.
NS	Quality Control failed: Not Spiked.

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number:		2509365-01				
Sample Name:		JW-1902-W02				
Date/Time Collected:		9/15/25 15:20				
Sample Matrix:		Water				
<u>Semivolatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>
Dibenz[a,h]anthracene	ug/L	< 5.00	E21	9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
1,2,4-Trichlorobenzene	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
1,2-Dichlorobenzene	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
1,3-Dichlorobenzene	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
1,4-Dichlorobenzene	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
1-Methylnaphthalene	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2,4,5-Trichlorophenol	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2,4,6-Trichlorophenol	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2,4-Dichlorophenol	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2,4-Dimethylphenol	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2,4-Dinitrophenol	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2,4-Dinitrotoluene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2,6-Dinitrotoluene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2-Chloronaphthalene	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2-Chlorophenol	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2-Methylnaphthalene	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2-Methylphenol	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2-Nitroaniline	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2-Nitrophenol	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2,2'-Oxybis(1-Chloropropane)	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
3 & 4-Methylphenol	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
3,3-Dichlorobenzidine	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
3-Nitroaniline	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
4,6-Dinitro-2-methylphenol	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
4-Bromophenyl-phenylether	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
4-Chloro-3-methylphenol	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
4-Chloroaniline	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
4-Chlorophenyl-phenylether	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
4-Nitroaniline	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
4-Nitrophenol	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Acenaphthene	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Acenaphthylene	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Aniline	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Anthracene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Azobenzene	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Benzidine	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Benzo (a) anthracene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Benzo[a]pyrene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Benzo[b]fluoranthene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Benzo[g,h,i]perylene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Benzo[k]fluoranthene	ug/L	< 5.00	9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018	
Benzoic acid	ug/L	< 15.0	9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018	
Benzyl alcohol	ug/L	< 5.00	9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018	

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number:		2509365-01				
Sample Name:		JW-1902-W02				
Date/Time Collected:		9/15/25 15:20				
Sample Matrix:		Water				
Semivolatiles	Units	Result	Qualifier(s)	Date/Time Analyzed	Batch	Method
Bis(2-chloroethoxy)methane	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Bis(2-chloroethyl)ether	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Bis(2-ethylhexyl)phthalate	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Butylbenzylphthalate	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Carbazole	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Chrysene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Dibenzofuran	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Diethylphthalate	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Dimethylphthalate	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Di-n-butylphthalate	ug/L	6.83		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Di-n-octylphthalate	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Fluoranthene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Fluorene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Hexachlorobenzene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Hexachlorobutadiene	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Hexachlorocyclopentadiene	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Hexachloroethane	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Indeno[1,2,3-cd]pyrene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Isophorone	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Naphthalene	ug/L	< 10.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Nitrobenzene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
N-Nitrosodimethylamine	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
N-Nitroso-di-n-propylamine	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
N-Nitrosodiphenylamine/diphenylamine	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Pentachlorophenol	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Phenanthrene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Phenol	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Pyrene	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Pyridine	ug/L	< 5.00		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2,4,6-Tribromophenol [surr]	%	87.7		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2-Fluorobiphenyl [surr]	%	69.3		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
2-Fluorophenol [surr]	%	50.0		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Nitrobenzene-d5 [surr]	%	65.6		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Phenol-d5 [surr]	%	38.4		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Terphenyl-d14 [surr]	%	132		9/17/25 15:45	B509326	SW 8270E, Rev. 6, 2018
Volatiles	Units	Result	Qualifier(s)	Date/Time Analyzed	Batch	Method
1,1,1,2-Tetrachloroethane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,1,1-Trichloroethane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,1,2,2-Tetrachloroethane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,1,2-Trichloroethane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,1-Dichloroethane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,1-Dichloroethene	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number:		2509365-01				
Sample Name:		JW-1902-W02				
Date/Time Collected:		9/15/25 15:20				
Sample Matrix:		Water				
<u>Volatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>
1,1-Dichloropropene	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,2,3-Trichlorobenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,2,3-Trichloropropane	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,2,4- Trimethylbenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,2,4-Trichlorobenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,2-Dibromo-3-chloropropane	ug/L	< 3.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,2-Dibromoethane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,2-Dichlorobenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,2-Dichloroethane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,2-Dichloropropane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,3,5- Trimethylbenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,3-Dichlorobenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,3-Dichloropropane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,4-Dichlorobenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
2,2-Dichloropropane	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
2-Butanone	ug/L	< 2.00	E-01	9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
2-Chloroethyl Vinyl Ether	ug/L	< 1.00	E20, E21, E2-	9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
2-Chlorotoluene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
2-Hexanone	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
4-Chlorotoluene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
4-Methyl-2-pentanone	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Acrolein	ug/L	< 4.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Acrylonitrile	ug/L	< 2.00	E-01	9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Benzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Bromobenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Bromochloromethane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Bromodichloromethane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Bromoform	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Bromomethane	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Carbon disulfide	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Carbon Tetrachloride	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Chlorobenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Dibromochloromethane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Chloroethane	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Chloroform	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Chloromethane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
cis-1,2-Dichloroethene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
cis-1,3-Dichloropropene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Dibromomethane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Dichlorodifluoromethane	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Ethylbenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Hexachlorobutadiene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Isopropylbenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006

24 September 2025



Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number: 2509365-01
Sample Name: JW-1902-W02
Date/Time Collected: 9/15/25 15:20
Sample Matrix: Water

<u>Volatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>
Methylene Chloride	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Methyl-tert-Butyl Ether	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Naphthalene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
m,p-Xylene	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
n-Butylbenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
n-Propylbenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
o-Xylene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
p-Isopropyltoluene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
sec-Butylbenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Styrene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
tert-Butylbenzene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Tetrachloroethene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Toluene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
trans-1,2-Dichloroethene	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
trans-1,3-Dichloropropene	ug/L	< 1.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Trichloroethene	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Trichlorofluoromethane	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Vinyl chloride	ug/L	< 2.00		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
4-Bromofluorobenzene [surr]	%	101		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
1,2-Dichloroethane-d4 [surr]	%	110		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006
Toluene-d8 [surr]	%	101		9/18/25 13:07	B509368	SW 8260C, Rev 3, 2006

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number: 2509365-02
Sample Name: JN-1278-W02
Date/Time Collected: 9/15/25 16:10
Sample Matrix: Water

<u>Semivolatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>
Dibenz[a,h]anthracene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
1,2,4-Trichlorobenzene	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
1,2-Dichlorobenzene	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
1,3-Dichlorobenzene	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
1,4-Dichlorobenzene	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
1-Methylnaphthalene	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
2,4,5-Trichlorophenol	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
2,4,6-Trichlorophenol	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
2,4-Dichlorophenol	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
2,4-Dimethylphenol	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
2,4-Dinitrophenol	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
2,4-Dinitrotoluene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
2,6-Dinitrotoluene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
2-Chloronaphthalene	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
2-Chlorophenol	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
2-Methylnaphthalene	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
2-Methylphenol	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
2-Nitroaniline	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
2-Nitrophenol	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
2,2'-Oxybis(1-Chloropropane)	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
3 & 4-Methylphenol	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
3,3-Dichlorobenzidine	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
3-Nitroaniline	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
4,6-Dinitro-2-methylphenol	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
4-Bromophenyl-phenylether	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
4-Chloro-3-methylphenol	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
4-Chloroaniline	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
4-Chlorophenyl-phenylether	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
4-Nitroaniline	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
4-Nitrophenol	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
Acenaphthene	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
Acenaphthylene	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
Aniline	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
Anthracene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
Azobenzene	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
Benzidine	ug/L	< 10.0	E21	9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
Benzo (a) anthracene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
Benzo[a]pyrene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
Benzo[b]fluoranthene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
Benzo[g,h,i]perylene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
Benzo[k]fluoranthene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
Benzoic acid	ug/L	< 15.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018
Benzyl alcohol	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number:		2509365-02					
Sample Name:		JN-1278-W02					
Date/Time Collected:		9/15/25 16:10					
Sample Matrix:		Water					
<u>Semivolatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>	
Bis(2-chloroethoxy)methane	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Bis(2-chloroethyl)ether	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Bis(2-ethylhexyl)phthalate	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Butylbenzylphthalate	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Carbazole	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Chrysene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Dibenzofuran	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Diethylphthalate	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Dimethylphthalate	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Di-n-butylphthalate	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Di-n-octylphthalate	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Fluoranthene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Fluorene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Hexachlorobenzene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Hexachlorobutadiene	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Hexachlorocyclopentadiene	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Hexachloroethane	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Indeno[1,2,3-cd]pyrene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Isophorone	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Naphthalene	ug/L	< 10.0		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Nitrobenzene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
N-Nitrosodimethylamine	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
N-Nitroso-di-n-propylamine	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
N-Nitrosodiphenylamine/diphenylamine	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Pentachlorophenol	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Phenanthrene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Phenol	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Pyrene	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Pyridine	ug/L	< 5.00		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
2,4,6-Tribromophenol [surr]	%	83.7		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
2-Fluorobiphenyl [surr]	%	75.2		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
2-Fluorophenol [surr]	%	54.5		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Nitrobenzene-d5 [surr]	%	71.3		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Phenol-d5 [surr]	%	43.8		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
Terphenyl-d14 [surr]	%	127		9/17/25 16:08	B509326	SW 8270E, Rev. 6, 2018	
<u>Volatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>	
1,1,1,2-Tetrachloroethane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006	
1,1,1-Trichloroethane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006	
1,1,2,2-Tetrachloroethane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006	
1,1,2-Trichloroethane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006	
1,1-Dichloroethane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006	
1,1-Dichloroethene	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006	
1,1-Dichloropropene	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006	

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number:		2509365-02				
Sample Name:		JN-1278-W02				
Date/Time Collected:		9/15/25 16:10				
Sample Matrix:		Water				
Volatiles	Units	Result	Qualifier(s)	Date/Time Analyzed	Batch	Method
1,2,3-Trichlorobenzene	ug/L	< 1.00	E-01 E21, E2-F	9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
1,2,3-Trichloropropane	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
1,2,4- Trimethylbenzene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
1,2,4-Trichlorobenzene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
1,2-Dibromo-3-chloropropane	ug/L	< 3.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
1,2-Dibromoethane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
1,2-Dichlorobenzene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
1,2-Dichloroethane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
1,2-Dichloropropane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
1,3,5- Trimethylbenzene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
1,3-Dichlorobenzene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
1,3-Dichloropropane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
1,4-Dichlorobenzene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
2,2-Dichloropropane	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
2-Butanone	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
2-Chloroethyl Vinyl Ether	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
2-Chlorotoluene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
2-Hexanone	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
4-Chlorotoluene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
4-Methyl-2-pentanone	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Acrolein	ug/L	< 4.00	E-01	9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Acrylonitrile	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Benzene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Bromobenzene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Bromochloromethane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Bromodichloromethane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Bromoform	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Bromomethane	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Carbon disulfide	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Carbon Tetrachloride	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Chlorobenzene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Dibromochloromethane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Chloroethane	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Chloroform	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Chloromethane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
cis-1,2-Dichloroethene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
cis-1,3-Dichloropropene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Dibromomethane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Dichlorodifluoromethane	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Ethylbenzene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Hexachlorobutadiene	ug/L	< 1.00	9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006	
Isopropylbenzene	ug/L	< 1.00	9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006	
Methylene Chloride	ug/L	< 2.00	9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006	

24 September 2025



Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number: 2509365-02
Sample Name: JN-1278-W02
Date/Time Collected: 9/15/25 16:10
Sample Matrix: Water

<u>Volatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>
Methyl-tert-Butyl Ether	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Naphthalene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
m,p-Xylene	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
n-Butylbenzene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
n-Propylbenzene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
o-Xylene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
p-Isopropyltoluene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
sec-Butylbenzene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Styrene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
tert-Butylbenzene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Tetrachloroethene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Toluene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
trans-1,2-Dichloroethene	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
trans-1,3-Dichloropropene	ug/L	< 1.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Trichloroethene	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Trichlorofluoromethane	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Vinyl chloride	ug/L	< 2.00		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
4-Bromofluorobenzene [surr]	%	102		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
1,2-Dichloroethane-d4 [surr]	%	111		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006
Toluene-d8 [surr]	%	102		9/18/25 13:31	B509368	SW 8260C, Rev 3, 2006

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number:		2509365-03					
Sample Name:		NBD-1114-W01					
Date/Time Collected:		9/15/25 17:15					
Sample Matrix:		Water					
<u>Semivolatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>	
Dibenz[a,h]anthracene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
1,2,4-Trichlorobenzene	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
1,2-Dichlorobenzene	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
1,3-Dichlorobenzene	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
1,4-Dichlorobenzene	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
1-Methylnaphthalene	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2,4,5-Trichlorophenol	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2,4,6-Trichlorophenol	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2,4-Dichlorophenol	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2,4-Dimethylphenol	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2,4-Dinitrophenol	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2,4-Dinitrotoluene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2,6-Dinitrotoluene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2-Chloronaphthalene	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2-Chlorophenol	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2-Methylnaphthalene	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2-Methylphenol	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2-Nitroaniline	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2-Nitrophenol	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2,2'-Oxybis(1-Chloropropane)	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
3 & 4-Methylphenol	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
3,3-Dichlorobenzidine	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
3-Nitroaniline	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
4,6-Dinitro-2-methylphenol	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
4-Bromophenyl-phenylether	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
4-Chloro-3-methylphenol	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
4-Chloroaniline	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
4-Chlorophenyl-phenylether	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
4-Nitroaniline	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
4-Nitrophenol	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Acenaphthene	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Acenaphthylene	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Aniline	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Anthracene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Azobenzene	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Benzidine	ug/L	< 10.0	E21	9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Benzo (a) anthracene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Benzo[a]pyrene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Benzo[b]fluoranthene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Benzo[g,h,i]perylene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Benzo[k]fluoranthene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Benzoic acid	ug/L	< 15.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Benzyl alcohol	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number:		2509365-03					
Sample Name:		NBD-1114-W01					
Date/Time Collected:		9/15/25 17:15					
Sample Matrix:		Water					
<u>Semivolatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>	
Bis(2-chloroethoxy)methane	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Bis(2-chloroethyl)ether	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Bis(2-ethylhexyl)phthalate	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Butylbenzylphthalate	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Carbazole	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Chrysene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Dibenzofuran	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Diethylphthalate	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Dimethylphthalate	ug/L	3440		9/17/25 19:11	B509326	SW 8270E, Rev. 6, 2018	
Di-n-butylphthalate	ug/L	38.9		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Di-n-octylphthalate	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Fluoranthene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Fluorene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Hexachlorobenzene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Hexachlorobutadiene	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Hexachlorocyclopentadiene	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Hexachloroethane	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Indeno[1,2,3-cd]pyrene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Isophorone	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Naphthalene	ug/L	< 10.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Nitrobenzene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
N-Nitrosodimethylamine	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
N-Nitroso-di-n-propylamine	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
N-Nitrosodiphenylamine/diphenylamine	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Pentachlorophenol	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Phenanthrene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Phenol	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Pyrene	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Pyridine	ug/L	< 5.00		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2,4,6-Tribromophenol [surr]	%	96.5		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2-Fluorobiphenyl [surr]	%	88.3		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
2-Fluorophenol [surr]	%	58.0		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Nitrobenzene-d5 [surr]	%	81.2		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Phenol-d5 [surr]	%	47.8		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
Terphenyl-d14 [surr]	%	136		9/17/25 16:31	B509326	SW 8270E, Rev. 6, 2018	
<u>Volatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>	
1,1,1,2-Tetrachloroethane	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006	
1,1,1-Trichloroethane	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006	
1,1,2,2-Tetrachloroethane	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006	
1,1,2-Trichloroethane	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006	
1,1-Dichloroethane	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006	
1,1-Dichloroethene	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006	

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number:		2509365-03				
Sample Name:		NBD-1114-W01				
Date/Time Collected:		9/15/25 17:15				
Sample Matrix:		Water				
<u>Volatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>
1,1-Dichloropropene	ug/L	< 2.00	E-01 E21, E2-F	9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
1,2,3-Trichlorobenzene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
1,2,3-Trichloropropane	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
1,2,4- Trimethylbenzene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
1,2,4-Trichlorobenzene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
1,2-Dibromo-3-chloropropane	ug/L	< 3.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
1,2-Dibromoethane	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
1,2-Dichlorobenzene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
1,2-Dichloroethane	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
1,2-Dichloropropane	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
1,3,5- Trimethylbenzene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
1,3-Dichlorobenzene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
1,3-Dichloropropane	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
1,4-Dichlorobenzene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
2,2-Dichloropropane	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
2-Butanone	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
2-Chloroethyl Vinyl Ether	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
2-Chlorotoluene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
2-Hexanone	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
4-Chlorotoluene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
4-Methyl-2-pentanone	ug/L	< 1.00	9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006	
Acrolein	ug/L	< 4.00	E-01	9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Acrylonitrile	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Benzene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Bromobenzene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Bromochloromethane	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Bromodichloromethane	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Bromoform	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Bromomethane	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Carbon disulfide	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Carbon Tetrachloride	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Chlorobenzene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Dibromochloromethane	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Chloroethane	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Chloroform	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Chloromethane	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
cis-1,2-Dichloroethene	ug/L	< 1.00	9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006	
cis-1,3-Dichloropropene	ug/L	< 1.00	9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006	
Dibromomethane	ug/L	< 1.00	9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006	
Dichlorodifluoromethane	ug/L	< 1.00	9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006	
Ethylbenzene	ug/L	< 1.00	9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006	
Hexachlorobutadiene	ug/L	< 1.00	9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006	
Isopropylbenzene	ug/L	< 1.00	9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006	

24 September 2025



Camille Gernhart
OSG
5301 Northshore Dr
North Little Rock, AR 72118
Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number: 2509365-03
Sample Name: NBD-1114-W01
Date/Time Collected: 9/15/25 17:15
Sample Matrix: Water

<u>Volatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>
Methylene Chloride	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Methyl-tert-Butyl Ether	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Naphthalene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
m,p-Xylene	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
n-Butylbenzene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
n-Propylbenzene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
o-Xylene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
p-Isopropyltoluene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
sec-Butylbenzene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Styrene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
tert-Butylbenzene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Tetrachloroethene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Toluene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
trans-1,2-Dichloroethene	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
trans-1,3-Dichloropropene	ug/L	< 1.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Trichloroethene	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Trichlorofluoromethane	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Vinyl chloride	ug/L	< 2.00		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
4-Bromofluorobenzene [surr]	%	98.2		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
1,2-Dichloroethane-d4 [surr]	%	112		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006
Toluene-d8 [surr]	%	104		9/18/25 13:56	B509368	SW 8260C, Rev 3, 2006

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number:		2509365-04				
Sample Name:		HM-SPR05				
Date/Time Collected:		9/16/25 9:25				
Sample Matrix:		Water				
<u>Semivolatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>
Dibenz[a,h]anthracene	ug/L	< 5.00	E21	9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
1,2,4-Trichlorobenzene	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
1,2-Dichlorobenzene	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
1,3-Dichlorobenzene	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
1,4-Dichlorobenzene	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
1-Methylnaphthalene	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
2,4,5-Trichlorophenol	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
2,4,6-Trichlorophenol	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
2,4-Dichlorophenol	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
2,4-Dimethylphenol	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
2,4-Dinitrophenol	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
2,4-Dinitrotoluene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
2,6-Dinitrotoluene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
2-Chloronaphthalene	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
2-Chlorophenol	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
2-Methylnaphthalene	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
2-Methylphenol	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
2-Nitroaniline	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
2-Nitrophenol	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
2,2'-Oxybis(1-Chloropropane)	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
3 & 4-Methylphenol	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
3,3-Dichlorobenzidine	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
3-Nitroaniline	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
4,6-Dinitro-2-methylphenol	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
4-Bromophenyl-phenylether	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
4-Chloro-3-methylphenol	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
4-Chloroaniline	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
4-Chlorophenyl-phenylether	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
4-Nitroaniline	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
4-Nitrophenol	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
Acenaphthene	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
Acenaphthylene	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
Aniline	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
Anthracene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
Azobenzene	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
Benzidine	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
Benzo (a) anthracene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
Benzo[a]pyrene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
Benzo[b]fluoranthene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
Benzo[g,h,i]perylene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018
Benzo[k]fluoranthene	ug/L	< 5.00	9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Benzoic acid	ug/L	< 15.0	9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Benzyl alcohol	ug/L	< 5.00	9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number:		2509365-04					
Sample Name:		HM-SPR05					
Date/Time Collected:		9/16/25 9:25					
Sample Matrix:		Water					
<u>Semivolatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>	
Bis(2-chloroethoxy)methane	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Bis(2-chloroethyl)ether	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Bis(2-ethylhexyl)phthalate	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Butylbenzylphthalate	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Carbazole	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Chrysene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Dibenzofuran	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Diethylphthalate	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Dimethylphthalate	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Di-n-butylphthalate	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Di-n-octylphthalate	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Fluoranthene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Fluorene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Hexachlorobenzene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Hexachlorobutadiene	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Hexachlorocyclopentadiene	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Hexachloroethane	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Indeno[1,2,3-cd]pyrene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Isophorone	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Naphthalene	ug/L	< 10.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Nitrobenzene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
N-Nitrosodimethylamine	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
N-Nitroso-di-n-propylamine	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
N-Nitrosodiphenylamine/diphenylamine	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Pentachlorophenol	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Phenanthrene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Phenol	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Pyrene	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Pyridine	ug/L	< 5.00		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
2,4,6-Tribromophenol [surr]	%	88.1		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
2-Fluorobiphenyl [surr]	%	84.4		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
2-Fluorophenol [surr]	%	62.3		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Nitrobenzene-d5 [surr]	%	85.0		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Phenol-d5 [surr]	%	47.9		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
Terphenyl-d14 [surr]	%	140		9/17/25 16:54	B509326	SW 8270E, Rev. 6, 2018	
<u>Volatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>	
1,1,1,2-Tetrachloroethane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006	
1,1,1-Trichloroethane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006	
1,1,2,2-Tetrachloroethane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006	
1,1,2-Trichloroethane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006	
1,1-Dichloroethane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006	
1,1-Dichloroethene	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006	
1,1-Dichloropropene	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006	

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number:		2509365-04				
Sample Name:		HM-SPR05				
Date/Time Collected:		9/16/25 9:25				
Sample Matrix:		Water				
<u>Volatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>
1,2,3-Trichlorobenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
1,2,3-Trichloropropane	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
1,2,4- Trimethylbenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
1,2,4-Trichlorobenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
1,2-Dibromo-3-chloropropane	ug/L	< 3.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
1,2-Dibromoethane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
1,2-Dichlorobenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
1,2-Dichloroethane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
1,2-Dichloropropane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
1,3,5- Trimethylbenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
1,3-Dichlorobenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
1,3-Dichloropropane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
1,4-Dichlorobenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
2,2-Dichloropropane	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
2-Butanone	ug/L	< 2.00	E-01	9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
2-Chloroethyl Vinyl Ether	ug/L	< 1.00	E21, E2-F	9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
2-Chlorotoluene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
2-Hexanone	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
4-Chlorotoluene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
4-Methyl-2-pentanone	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Acrolein	ug/L	< 4.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Acrylonitrile	ug/L	< 2.00	E-01	9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Benzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Bromobenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Bromochloromethane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Bromodichloromethane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Bromoform	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Bromomethane	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Carbon disulfide	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Carbon Tetrachloride	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Chlorobenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Dibromochloromethane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Chloroethane	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Chloroform	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Chloromethane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
cis-1,2-Dichloroethene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
cis-1,3-Dichloropropene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Dibromomethane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Dichlorodifluoromethane	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Ethylbenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Hexachlorobutadiene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Isopropylbenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Methylene Chloride	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006

24 September 2025



Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

ANALYTICAL RESULTS

Lab Number: 2509365-04
Sample Name: HM-SPR05
Date/Time Collected: 9/16/25 9:25
Sample Matrix: Water

<u>Volatiles</u>	<u>Units</u>	<u>Result</u>	<u>Qualifier(s)</u>	<u>Date/Time Analyzed</u>	<u>Batch</u>	<u>Method</u>
Methyl-tert-Butyl Ether	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Naphthalene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
m,p-Xylene	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
n-Butylbenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
n-Propylbenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
o-Xylene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
p-Isopropyltoluene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
sec-Butylbenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Styrene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
tert-Butylbenzene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Tetrachloroethene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Toluene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
trans-1,2-Dichloroethene	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
trans-1,3-Dichloropropene	ug/L	< 1.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Trichloroethene	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Trichlorofluoromethane	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Vinyl chloride	ug/L	< 2.00		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
4-Bromofluorobenzene [surr]	%	97.9		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
1,2-Dichloroethane-d4 [surr]	%	111		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006
Toluene-d8 [surr]	%	103		9/18/25 14:21	B509368	SW 8260C, Rev 3, 2006

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

QUALITY CONTROL RESULTS

Semivolatiles -- Batch: B509326 (Water)

Prepared: 17-Sep-25 10:32 By: JAH -- Analyzed: 17-Sep-25 15:22 By: TB

Analyte	BLK	LCS / LCSD	MS / MSD	Dup	RPD	Qualifiers
1,2,4-Trichlorobenzene	<10.0 ug/L	64.8% / NA	62.9% / 61.4%		2.37%	
1,2-Dichlorobenzene	<10.0 ug/L	66.2% / NA	63.1% / 62.8%		0.453%	
1,3-Dichlorobenzene	<10.0 ug/L	66.1% / NA	64.3% / 62.8%		2.28%	
1,4-Dichlorobenzene	<5.00 ug/L	70.3% / NA	66.5% / 66.0%		0.774%	
1-Methylnaphthalene	<10.0 ug/L	73.0% / NA	72.0% / 70.5%		2.17%	
2,2'-Oxybis(1-Chloropropane)	<10.0 ug/L	79.4% / NA	77.6% / 77.3%		0.443%	
2,4,5-Trichlorophenol	<5.00 ug/L	92.3% / NA	94.0% / 92.3%		1.85%	
2,4,6-Trichlorophenol	<5.00 ug/L	91.6% / NA	91.5% / 88.1%		3.77%	
2,4-Dichlorophenol	<5.00 ug/L	69.7% / NA	69.8% / 69.8%		0.00230%	
2,4-Dimethylphenol	<5.00 ug/L	64.6% / NA	63.7% / 62.8%		1.33%	
2,4-Dinitrophenol	<10.0 ug/L	101% / NA	97.6% / 100%		2.82%	
2,4-Dinitrotoluene	<5.00 ug/L	105% / NA	106% / 103%		2.63%	
2,6-Dinitrotoluene	<5.00 ug/L	97.3% / NA	97.1% / 95.5%		1.69%	
2-Chloronaphthalene	<10.0 ug/L	84.2% / NA	82.5% / 81.3%		1.44%	
2-Chlorophenol	<5.00 ug/L	76.9% / NA	77.0% / 75.6%		1.84%	
2-Methylnaphthalene	<10.0 ug/L	72.1% / NA	69.8% / 69.5%		0.416%	
2-Methylphenol	<5.00 ug/L	79.0% / NA	78.5% / 79.0%		0.549%	
2-Nitroaniline	<10.0 ug/L	90.8% / NA	91.8% / 88.8%		3.38%	
2-Nitrophenol	<5.00 ug/L	82.2% / NA	81.2% / 81.3%		0.167%	
3 & 4-Methylphenol	<5.00 ug/L	77.2% / NA	76.4% / 77.0%		0.842%	
3,3-Dichlorobenzidine	<5.00 ug/L	77.1% / NA	77.8% / 74.8%		3.81%	
3-Nitroaniline	<5.00 ug/L	83.6% / NA	86.3% / 85.2%		1.34%	
4,6-Dinitro-2-methylphenol	<10.0 ug/L	93.5% / NA	100% / 94.8%		5.58%	
4-Bromophenyl-phenylether	<5.00 ug/L	88.2% / NA	90.5% / 83.7%		7.80%	
4-Chloro-3-methylphenol	<5.00 ug/L	78.5% / NA	79.1% / 77.6%		1.87%	
4-Chloroaniline	<5.00 ug/L	65.2% / NA	65.4% / 65.4%		0.00908%	
4-Chlorophenyl-phenylether	<10.0 ug/L	95.8% / NA	95.5% / 92.4%		3.30%	
4-Nitroaniline	<5.00 ug/L	94.6% / NA	94.6% / 93.3%		1.42%	
4-Nitrophenol	<5.00 ug/L	78.0% / NA	75.4% / 75.5%		0.0977%	
Acenaphthene	<10.0 ug/L	90.2% / NA	90.0% / 87.6%		2.79%	
Acenaphthylene	<10.0 ug/L	92.1% / NA	90.8% / 89.5%		1.51%	
Aniline	<5.00 ug/L	57.4% / NA	57.3% / 57.5%		0.311%	
Anthracene	<5.00 ug/L	89.5% / NA	91.3% / 88.5%		3.08%	
Azobenzene	<10.0 ug/L	91.0% / NA	91.9% / 89.7%		2.42%	
Benzidine	<10.0 ug/L	11.3% / NA	11.0% / 11.2%		2.09%	E21
Benzo (a) anthracene	<5.00 ug/L	90.2% / NA	93.4% / 90.4%		3.26%	
Benzo[a]pyrene	<5.00 ug/L	98.5% / NA	97.2% / 95.3%		2.02%	
Benzo[b]fluoranthene	<5.00 ug/L	100% / NA	97.4% / 96.7%		0.751%	
Benzo[g,h,i]perylene	<5.00 ug/L	94.0% / NA	92.7% / 91.8%		0.946%	
Benzo[k]fluoranthene	<5.00 ug/L	98.3% / NA	97.4% / 96.5%		0.991%	
Benzoic acid	<15.0 ug/L	59.4% / NA	65.2% / 65.9%		1.07%	
Benzyl alcohol	<5.00 ug/L	80.4% / NA	79.9% / 79.7%		0.191%	
Bis(2-chloroethoxy)methane	<5.00 ug/L	74.0% / NA	73.9% / 72.9%		1.41%	
Bis(2-chloroethyl)ether	<5.00 ug/L	73.7% / NA	72.5% / 71.8%		0.964%	
Bis(2-ethylhexyl)phthalate	<5.00 ug/L	94.2% / NA	98.3% / 93.3%		5.18%	
Butylbenzylphthalate	<5.00 ug/L	98.1% / NA	98.8% / 95.2%		3.67%	
Carbazole	<5.00 ug/L	86.9% / NA	86.2% / 87.4%		1.29%	
Chrysene	<5.00 ug/L	89.0% / NA	90.1% / 86.4%		4.25%	

Camille Gernhart
OSG
5301 Northshore Dr
North Little Rock, AR 72118
Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

QUALITY CONTROL RESULTS

Semivolatiles -- Batch: B509326 (Water)

Prepared: 17-Sep-25 10:32 By: JAH -- Analyzed: 17-Sep-25 15:22 By: TB

Analyte	BLK	LCS / LCSD	MS / MSD	Dup	RPD	Qualifiers
Dibenz[a,h]anthracene	<5.00 ug/L	50.6% / NA	50.3% / 49.0%		2.49%	
Dibenzofuran	<5.00 ug/L	88.6% / NA	88.6% / 86.8%		2.09%	
Diethylphthalate	<5.00 ug/L	98.4% / NA	98.5% / 95.9%		2.63%	
Dimethylphthalate	<5.00 ug/L	95.9% / NA	97.3% / 93.4%		4.09%	
Di-n-butylphthalate	<5.00 ug/L	96.7% / NA	98.1% / 96.6%		1.43%	
Di-n-octylphthalate	<5.00 ug/L	101% / NA	103% / 99.0%		4.15%	
Fluoranthene	<5.00 ug/L	91.6% / NA	93.5% / 88.9%		5.06%	
Fluorene	<5.00 ug/L	93.5% / NA	93.3% / 91.2%		2.26%	
Hexachlorobenzene	<5.00 ug/L	86.3% / NA	84.7% / 82.0%		3.22%	
Hexachlorobutadiene	<10.0 ug/L	66.2% / NA	63.3% / 63.6%		0.577%	
Hexachlorocyclopentadiene	<10.0 ug/L	73.5% / NA	75.9% / 73.9%		2.67%	
Hexachloroethane	<5.00 ug/L	71.5% / NA	68.3% / 67.3%		1.50%	
Indeno[1,2,3-cd]pyrene	<5.00 ug/L	98.4% / NA	97.2% / 94.9%		2.36%	
Isophorone	<10.0 ug/L	89.1% / NA	90.0% / 91.0%		1.13%	
Naphthalene	<10.0 ug/L	71.1% / NA	69.5% / 68.7%		1.22%	
Nitrobenzene	<5.00 ug/L	78.7% / NA	75.6% / 76.7%		1.43%	
N-Nitrosodimethylamine	<5.00 ug/L	52.6% / NA	53.1% / 55.5%		4.29%	
N-Nitroso-di-n-propylamine	<5.00 ug/L	84.9% / NA	82.4% / 82.4%		0.0356%	
N-Nitrosodiphenylamine/diphenylamine	<5.00 ug/L	88.7% / NA	88.9% / 87.4%		1.76%	
Pentachlorophenol	<4.00 ug/L	101% / NA	103% / 96.7%		5.83%	
Phenanthrene	<5.00 ug/L	91.7% / NA	92.1% / 89.3%		3.12%	
Phenol	<5.00 ug/L	47.8% / NA	47.8% / 49.1%		2.77%	
Pyrene	<5.00 ug/L	90.1% / NA	94.6% / 90.8%		4.01%	
Pyridine	<5.00 ug/L	39.2% / NA	41.1% / 42.8%		3.94%	
2,4,6-Tribromophenol [surr]	95.1 %	100% / NA	101% / 100%		NA	
2-Fluorobiphenyl [surr]	85.1 %	93.9% / NA	89.7% / 89.5%		NA	
2-Fluorophenol [surr]	62.0 %	63.6% / NA	60.7% / 62.1%		NA	
Nitrobenzene-d5 [surr]	83.9 %	87.3% / NA	83.8% / 82.8%		NA	
Phenol-d5 [surr]	48.1 %	50.9% / NA	50.0% / 51.6%		NA	
Terphenyl-d14 [surr]	138 %	129% / NA	131% / 126%		NA	

Camille Gernhart

OSG

5301 Northshore Dr

North Little Rock, AR 72118

Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05

QUALITY CONTROL RESULTS**Volatiles -- Batch: B509368 (Water)**

Prepared: 18-Sep-25 10:38 By: TB -- Analyzed: 18-Sep-25 15:10 By: tb

Analyte	BLK	LCS / LCSD		MS / MSD		Dup	RPD	Qualifiers
1,1,1,2-Tetrachloroethane	<1.00 ug/L	94.5%	/ NA	95.5%	/ 93.2%		2.48%	
1,1,1-Trichloroethane	<1.00 ug/L	94.9%	/ NA	103%	/ 102%		1.07%	
1,1,2,2-Tetrachloroethane	<1.00 ug/L	96.9%	/ NA	101%	/ 94.1%		6.97%	
1,1,2-Trichloroethane	<1.00 ug/L	93.5%	/ NA	97.2%	/ 94.5%		2.85%	
1,1-Dichloroethane	<1.00 ug/L	108%	/ NA	112%	/ 111%		0.915%	
1,1-Dichloroethene	<2.00 ug/L	97.6%	/ NA	106%	/ 105%		1.33%	
1,1-Dichloropropene	<2.00 ug/L	98.9%	/ NA	107%	/ 105%		1.87%	
1,2,3-Trichlorobenzene	<1.00 ug/L	91.8%	/ NA	95.7%	/ 90.3%		5.79%	
1,2,3-Trichloropropane	<2.00 ug/L	86.7%	/ NA	87.3%	/ 87.5%		0.309%	
1,2,4- Trimethylbenzene	<1.00 ug/L	87.9%	/ NA	91.0%	/ 87.6%		3.87%	
1,2,4-Trichlorobenzene	<1.00 ug/L	88.6%	/ NA	91.2%	/ 88.5%		3.00%	
1,2-Dibromo-3-chloropropane	<3.00 ug/L	94.6%	/ NA	96.0%	/ 95.7%		0.334%	
1,2-Dibromoethane	<1.00 ug/L	92.2%	/ NA	92.3%	/ 93.4%		1.24%	
1,2-Dichlorobenzene	<1.00 ug/L	89.9%	/ NA	92.1%	/ 87.7%		4.99%	
1,2-Dichloroethane	<1.00 ug/L	100%	/ NA	101%	/ 98.8%		2.31%	
1,2-Dichloropropane	<1.00 ug/L	104%	/ NA	105%	/ 103%		1.44%	
1,3,5- Trimethylbenzene	<1.00 ug/L	88.1%	/ NA	91.1%	/ 88.1%		3.40%	
1,3-Dichlorobenzene	<1.00 ug/L	90.8%	/ NA	91.8%	/ 87.6%		4.62%	
1,3-Dichloropropane	<1.00 ug/L	97.7%	/ NA	100%	/ 98.9%		1.07%	
1,4-Dichlorobenzene	<1.00 ug/L	88.1%	/ NA	91.9%	/ 86.7%		5.91%	
2,2-Dichloropropane	<2.00 ug/L	98.0%	/ NA	106%	/ 103%		2.27%	
2-Butanone	<2.00 ug/L	126%	/ NA	123%	/ 123%		0.233%	E-01
2-Chloroethyl Vinyl Ether	<1.00 ug/L	59.6%	/ NA	MBI	/ MBI		%	E21, E2-F, MBI
2-Chlorotoluene	<1.00 ug/L	88.4%	/ NA	90.4%	/ 87.0%		3.81%	
2-Hexanone	<2.00 ug/L	118%	/ NA	120%	/ 120%		0.549%	
4-Chlorotoluene	<1.00 ug/L	88.7%	/ NA	91.9%	/ 86.8%		5.70%	
4-Methyl-2-pentanone	<1.00 ug/L	119%	/ NA	122%	/ 122%		0.140%	
Acrolein	<4.00 ug/L	105%	/ NA	108%	/ 110%		2.15%	
Acrylonitrile	<2.00 ug/L	120%	/ NA	118%	/ 122%		3.46%	E-01
Benzene	<1.00 ug/L	98.0%	/ NA	104%	/ 102%		1.63%	
Bromobenzene	<1.00 ug/L	87.2%	/ NA	87.2%	/ 85.5%		1.99%	
Bromochloromethane	<1.00 ug/L	93.5%	/ NA	100%	/ 100%		0.178%	
Bromodichloromethane	<1.00 ug/L	94.3%	/ NA	98.0%	/ 93.2%		5.01%	
Bromoform	<1.00 ug/L	95.7%	/ NA	92.8%	/ 93.2%		0.393%	
Bromomethane	<2.00 ug/L	103%	/ NA	108%	/ 105%		2.52%	
Carbon disulfide	<1.00 ug/L	86.8%	/ NA	94.5%	/ 92.7%		1.87%	
Carbon Tetrachloride	<2.00 ug/L	92.7%	/ NA	104%	/ 103%		0.953%	
Chlorobenzene	<1.00 ug/L	91.0%	/ NA	92.6%	/ 90.0%		2.86%	
Chloroethane	<2.00 ug/L	105%	/ NA	120%	/ 117%		2.82%	
Chloroform	<2.00 ug/L	101%	/ NA	107%	/ 103%		3.77%	
Chloromethane	<1.00 ug/L	110%	/ NA	116%	/ 114%		1.75%	
cis-1,2-Dichloroethene	<1.00 ug/L	100%	/ NA	107%	/ 108%		0.925%	
cis-1,3-Dichloropropene	<1.00 ug/L	91.0%	/ NA	92.9%	/ 91.3%		1.76%	
Dibromochloromethane	<1.00 ug/L	90.9%	/ NA	92.3%	/ 92.2%		0.114%	
Dibromomethane	<1.00 ug/L	91.2%	/ NA	96.8%	/ 91.6%		5.45%	
Dichlorodifluoromethane	<1.00 ug/L	98.1%	/ NA	109%	/ 107%		1.27%	
Ethylbenzene	<1.00 ug/L	92.7%	/ NA	97.4%	/ 94.1%		3.40%	
Hexachlorobutadiene	<1.00 ug/L	85.8%	/ NA	95.4%	/ 89.6%		6.23%	

24 September 2025

Camille Gernhart
OSG
5301 Northshore Dr
North Little Rock, AR 72118
Project: Samples

Project Number: September 2025 -- EVT

Date Received: 16-Sep-25 14:05



QUALITY CONTROL RESULTS

Volatiles -- Batch: B509368 (Water)

Prepared: 18-Sep-25 10:38 By: TB -- Analyzed: 18-Sep-25 15:10 By: tb

Analyte	BLK	LCS / LCSD	MS / MSD	Dup	RPD	Qualifiers
Isopropylbenzene	<1.00 ug/L	85.8% / NA	91.0% / 86.4%		5.22%	
m,p-Xylene	<2.00 ug/L	91.5% / NA	96.0% / 93.7%		2.40%	
Methylene Chloride	<2.00 ug/L	96.4% / NA	95.7% / 98.8%		3.26%	
Methyl-tert-Butyl Ether	<1.00 ug/L	105% / NA	109% / 108%		0.413%	
Naphthalene	<1.00 ug/L	82.7% / NA	85.0% / 80.9%		4.95%	
n-Butylbenzene	<1.00 ug/L	89.1% / NA	96.0% / 90.2%		6.17%	
n-Propylbenzene	<1.00 ug/L	86.7% / NA	92.8% / 88.8%		4.33%	
o-Xylene	<1.00 ug/L	90.6% / NA	93.8% / 91.3%		2.70%	
p-Isopropyltoluene	<1.00 ug/L	86.8% / NA	92.3% / 87.9%		4.89%	
sec-Butylbenzene	<1.00 ug/L	83.3% / NA	92.2% / 86.9%		5.84%	
Styrene	<1.00 ug/L	92.1% / NA	95.0% / 91.3%		4.02%	
tert-Butylbenzene	<1.00 ug/L	84.0% / NA	89.1% / 84.2%		5.73%	
Tetrachloroethene	<1.00 ug/L	90.9% / NA	97.0% / 96.4%		0.651%	
Toluene	<1.00 ug/L	94.2% / NA	99.3% / 96.4%		3.03%	
trans-1,2-Dichloroethene	<2.00 ug/L	99.2% / NA	106% / 104%		2.75%	
trans-1,3-Dichloropropene	<1.00 ug/L	91.4% / NA	92.0% / 91.3%		0.707%	
Trichloroethene	<2.00 ug/L	84.9% / NA	91.0% / 89.1%		2.07%	
Trichlorofluoromethane	<2.00 ug/L	95.9% / NA	107% / 105%		2.23%	
Vinyl chloride	<2.00 ug/L	102% / NA	110% / 108%		1.99%	
1,2-Dichloroethane-d4 [surr]	110 %	111% / NA	112% / 116%		NA	
4-Bromofluorobenzene [surr]	98.9 %	99.7% / NA	98.0% / 98.0%		NA	
Toluene-d8 [surr]	102 %	103% / NA	102% / 104%		NA	

QUALIFIER(S)

- *E-01: Estimated Result; This Analyte Failed "High" in the CCV; If the sample is non-detect for this analyte, the CCV demonstrated the analyte would have been detected were it present.
- *E20: Estimated Result Due to Matrix Spike and/or Matrix Spike Duplicate Failure; This sample was used as the "parent sample" in MS/MSD prep.
- *E21: Estimated Result; This Analyte failed (low) in the CCV.
- *E2-F: Second Source Verification Failure
- *MBI: Masked By Interference

All Analysis performed according to EPA approved methodology when available :

SW 846, Revised December, 1996; EPA 600/4-79-020, Revised March, 1983; Standard Methods.

Instrument calibration and quality control samples performed at or above frequency specified in analytical method.

Reviewed by: 
Norma James
Technical Director

[illegible]